



*Delaware Health
And Social Services*

DIVISION OF MANAGEMENT SERVICES

PROCUREMENT

DATE: July 29, 2010

RFP # HSS-10-092

PREFERRED DIABETIC SUPPLIES

for

DIVISION OF MEDICAID AND MEDICAL ASSISTANCE

Date Due: AUGUST 30, 2010

By 11:00 A.M. LOCAL TIME

ADDENDUM # 1 - TECHNICAL PROPOSAL

THE ATTACHED SHEETS HEREBY BECOME A PART OF THE ABOVE
MENTIONED RFP **HSS-10-092**.

PLEASE NOTE:

**REVISIONS TO TECHNICAL PROPOSAL FOUND ON PAGE 5 and
PAGE 6 OF THE RFP**

BRUCE KRUG
PROCUREMENT ADMINISTRATOR
(302) 255-9291

BETH LAUCIUS
Chief of Administration, DMMA
(302)255-9525

Four **changes** are being made in association with the clinical aspects found in the Technical Proposal appearing on Pages 5 and 6 of RFP **HSS-10-092**. These **changes** are:

1. (Page 5) Striking a.
2. (Page 5) Modifying c.
3. (Page 5) Striking h.
4. (Page 6) Adding sentence to first paragraph

A. Technical Proposal

....

Proposals must describe and meet each of the following clinical requirements:

- ~~a. Manufacturer/labeler (the bidder) must offer a minimum of 1 blood glucose monitor that is capable of testing for blood ketones on the same monitor.~~
- b. Manufacturer/labeler (the bidder) must offer a minimum of 1 monitor that offers alternate site testing and have at least 1 alternate site with FDA-approved testing equivalency to the finger.
- c. Manufacturer/labeler (the bidder) must offer a minimum of 1 monitor that requires a test strip sample size that is equal to or less than ~~0.3~~ **1.0** micro liters of blood and provides test results in less than 5 seconds.
- d. Manufacturer/labeler (the bidder) must offer a minimum of 1 monitor with a backlit screen for clients that have difficulty with their eyesight.
- e. Manufacturer/labeler (the bidder) must describe the size, weight, and portability of the monitor.
- f. Manufacturer/labeler (the bidder) must offer a minimum of 1 monitor with computer downloading capability with a minimum of 250 readings stored.
- g. Manufacturer/labeler (the bidder) must offer a minimum of 1 monitor that has automatic coding or check strip.
- ~~h. Manufacturer/labeler (the bidder) must provide monitors that are capable of detecting interfering substances in the blood sample to ensure accurate blood glucose readings.~~

The proposal must include complete details regarding products being offered in the proposal such as the quality and use of the product in meeting the medical service needs of the population identified in this RFP. **In particular, the bidder must address the performance accuracy of the product(s) in terms of blood glucose readings with attention to specific factors which affect such readings, for example, the ability to detect interfering substances in the blood sample.**

Samples of products...